

# **EXACTA** MICROPIPETTES



# EXACTA MICROPIPETTES



Pipette holder  
Code K7506

## MODELS

MODEL ( $\mu$ l)	CAPACITY ( $\mu$ l)		ACCURACY (%)	REPEATABILITY (%)
S - 2 Code K7507	0.2 ~ 2	at 0.2 $\mu$ l	$\pm$ 12.0%	$\leq$ 6.00%
		at 0.5 $\mu$ l	$\pm$ 5.0%	$\leq$ 2.50%
		at 2 $\mu$ l	$\pm$ 1.5%	$\leq$ 0.70%
S - 10 Code K7508	0.5 ~ 10	at 1 $\mu$ l	$\pm$ 2.5%	$\leq$ 1.25%
		at 5 $\mu$ l	$\pm$ 1.5%	$\leq$ 0.60%
		at 10 $\mu$ l	$\pm$ 1.0%	$\leq$ 0.40%
S - 20 Code K7501	2 ~ 20	at 5 $\mu$ l	$\pm$ 2.0%	$\leq$ 0.80%
		at 10 $\mu$ l	$\pm$ 1.0%	$\leq$ 0.50%
		at 20 $\mu$ l	$\pm$ 1.0%	$\leq$ 0.30%
S - 100 Code K7502	20 ~ 100	at 20 $\mu$ l	$\pm$ 1.8%	$\leq$ 0.50%
		at 50 $\mu$ l	$\pm$ 0.8%	$\leq$ 0.24%
		at 100 $\mu$ l	$\pm$ 0.8%	$\leq$ 0.15%
S - 200 Code K 7503	50 ~ 200	at 50 $\mu$ l	$\pm$ 1.0%	$\leq$ 0.40%
		at 100 $\mu$ l	$\pm$ 0.8%	$\leq$ 0.25%
		at 200 $\mu$ l	$\pm$ 0.8%	$\leq$ 0.15%
S - 1000 Code K7504	100 ~ 1000	at 200 $\mu$ l	$\pm$ 1.5%	$\leq$ 0.30%
		at 500 $\mu$ l	$\pm$ 0.8%	$\leq$ 0.20%
		at 1000 $\mu$ l	$\pm$ 0.8%	$\leq$ 0.15%
S - 5000 Code K7505	1000 ~ 5000	at 1000 $\mu$ l	$\pm$ 1.2%	$\leq$ 0.30%
		at 2500 $\mu$ l	$\pm$ 0.6%	$\leq$ 0.25%
		at 5000 $\mu$ l	$\pm$ 0.6%	$\leq$ 0.16%

## EXACTA MICROPIPETTES

The EXACTA variable volume micropipette range comprises high quality, manual instruments at very competitive prices. The pipettes operate using the air displacement principle, produced by the movement of an internal piston, thus avoiding possible contamination between aspirations. Each micropipette is supplied with a certificate, issued following gravimetric test checks, declaring the maximum capacity's mean value, the accuracy and repeatability.

## PIPETTE HOUSING

The pipette housing has been designed with a comfortable and ergonomic handgrip and a smooth running, uniform piston action; these features minimize mistakes due to fatigue or incorrect handling, even whilst wearing gloves. The construction materials used, characterized by excellent temperature resistance, allow the micropipettes to be sterilized in an autoclave.

## EASY TO USE

Volume setting can be achieved single-handedly, and an efficient O-ring system avoids accidental adjustment of the set volume. Finally, the close proximity of the plunger and tip ejection buttons enables extremely simple and rapid operation.

Authorized dealer:



GERMANY  
FRANCE  
ITALY  
SPAIN  
THE NETHERLANDS  
UNITED KINGDOM

**EXACTA + OPTECH**<sup>®</sup> GmbH

Lerchenstrasse 10, D-80995 München - Germany  
Tel. +49-89-157 80 661 Fax: +49-89-157 80 662  
Website: [www.exactaoptech.com](http://www.exactaoptech.com)